

Interpretation of Pulse Oximetry Results

	RIGHT HAND												
ONE FOOT	%	100	99	98	97	96	95	94	93	92	91	90	89
	100	PASS	PASS	PASS	PASS	REPEAT	FAIL						
	99	PASS	PASS	PASS	PASS	PASS	REPEAT	REPEAT	REPEAT	REPEAT	REPEAT	REPEAT	FAIL
	98	PASS	PASS	PASS	PASS	PASS	PASS	REPEAT	REPEAT	REPEAT	REPEAT	REPEAT	FAIL
	97	PASS	REPEAT	REPEAT	REPEAT	REPEAT	FAIL						
	96	REPEAT	PASS	REPEAT	REPEAT	REPEAT	FAIL						
	95	REPEAT	REPEAT	PASS	REPEAT	REPEAT	FAIL						
	94	REPEAT	REPEAT	REPEAT	PASS	PASS	PASS	REPEAT	REPEAT	REPEAT	REPEAT	REPEAT	FAIL
	93	REPEAT	REPEAT	REPEAT	REPEAT	PASS	PASS	REPEAT	REPEAT	REPEAT	REPEAT	REPEAT	FAIL
	92	REPEAT	REPEAT	REPEAT	REPEAT	REPEAT	PASS	REPEAT	REPEAT	REPEAT	REPEAT	REPEAT	FAIL
	91	REPEAT	FAIL										
	90	REPEAT	FAIL										
	89	FAIL	FAIL										

<u>Pulse Oximetry Screening Protocol & Interpretation Table</u> reproduced with permission from author:

Wong, K.K., Fournier, A., Fruitman, D.S., et al. (2017). Canadian Cardiovascular Society / Canadian Pediatric Cardiology Association position statement on pulse oximetry screening in newborns to enhance detection of critical congenital heart disease. Canadian Journal of Cardiology 33, 199-208.